



cewal UK LTD

• Pressure & Temperature Instrumentation & Controls Manufacturer •

Unit 4c Woodhall Business Park Sudbury Suffolk CO10 1WH United Kingdom

Tel: 01787 882038 Fax: 01787 882039 E-mail: info@cewal.co.uk Web: www.cewal.co.uk

POINT TO POINT CALIBRATION CERTIFICATE NO: 10099

TO : CUSTOMER NAME
CUSTOMER ADDRESS

INSTRUMENT SERIAL NO : C04855-8

OUR JOB NO : 4855

ITEM	QTY	INSTRUMENT DESCRIPTION WITH IDENTIFICATION/REF/PART CODE
1	1	100MM DIAMETER PRESSURE GAUGE, BLACK STEEL CASE, BRASS INTERNALS, BOTTOM ENTRY, 3/8 BSP THREAD, 0-40 BAR/PSI, CLASS 1.0, DRY PRODUCT CODE B401A356L

INSTRUMENT READING SCALED BAR	TEST PRESSURE READING RISING BAR	TEST PRESSURE READING FALLING BAR
0	-	-
5	4.93	4.85
10	9.89	9.78
15	14.92	14.78
20	19.80	19.72
25	24.71	24.66
30	29.68	29.67
35	34.72	34.65
40	39.78	-

We certify that the above instrument has been tested to BS EN 837-1 : 1998. Accuracy tested to and within $\pm 1.00\%$ FSD limits. Reading as shown above.

Instrument tested with: WATER medium
Instrument tested in: VERTICAL position
Laboratory Temperature $20^{\circ}\text{C} \pm 2^{\circ}\text{C}$

Druck DPI 602 Electronic Vacuum/Pressure Tester Serial No: 998/96-1
Transducer Serial No: 1240396

Calibration equipment above is within its period of calibration and traceable to NPL/National Standards

Signed:

Date:

