



# cewal UK LTD

• Pressure & Temperature Instrumentation & Controls Manufacturer •

Unit 4c Woodhall Business Park Sudbury Suffolk CO10 1WH United Kingdom

Tel: 01787 882038 Fax: 01787 882039 E-mail: info@cewal.co.uk Web: www.cewal.co.uk

## STANDARD TEST CERTIFICATE NO : 10121

TO : CUSTOMER NAME  
CUSTOMER ADDRESS

INSTRUMENT SERIAL NO : C04883

ITEM	QTY	INSTRUMENT DESCRIPTION WITH IDENTIFICATION/REF/PART CODE
1	1	100MM DIAMETER PRESSURE GAUGE STAINLESS STEEL CASE BRASS INTERNALS BOTTOM ENTRY 1/2" BSP THREAD 0-16 BAR/PSI CLASS 1.0 DRY  PRODUCT CODE 90493530

We certify that the above instrument has been tested to BS EN 837-1 : 1998. Accuracy tested to and within  $\pm 1.00\%$  FSD limits.

Instrument tested with: WATER medium

Instrument tested in: VERTICAL position

Laboratory Temperature  $20^{\circ}\text{C} \pm 2^{\circ}\text{C}$

Druck DPI 602 Electronic Vacuum/Pressure Tester Serial No: 998/96-1

Transducer Serial No: 1642706

Calibration equipment above is within its period of calibration and traceable to NPL/National Standards

Signed:

Date:

