

ACETYLENE USE GAUGES

Acetylene when in contact with high copper or silver content alloy may form an explosion risk. All parts of the gauge in contact with acetylene should be steel or 65% maximum copper content alloy. Acetylene gauges must be of the safety pattern type and have Acetylene printed clearly on the dial.

These instruments are built to special order to customer specification. Please contact us for further information or quotation.



cewal UK LTD

Unit 4C • Woodhall Business Park • Sudbury • Suffolk • CO10 1WH

Tel: +44 (0) 1787 882038 • Fax: +44 (0) 1787 882039

Email: info@cewal.co.uk • Website: www.cewal.co.uk